



## **ΜΙΧΑΗΛΙΔΗΣ Σ. & Ε. & ΣΙΑ Ο.Ε.**

ΑΘΗΝΑ : Περικλέους 22, ΤΚ-10562, Σύνταγμα,  
▲ Τηλ. 210/3242270, Fax. 210/3255457  
e-mail : [sales@michailidis.com.gr](mailto:sales@michailidis.com.gr)

Πάγκοι

# *osada success 40*



Our micromotor performance, keeping its excellence, changes the image of dental technique and jewelry processing.

- This space-efficient product saves desktop working space and provides a comfortable working environment, due to its silent rotation and small oscillations.
- The minimum rotation speed is 1,000 min<sup>-1</sup>! Excellent work may be accomplished at low speeds. A smooth surface is rendered by stable torque.
- A lever type brushless motor is adopted, enabling bur exchange, without stopping the handpiece motor.

Product name (Model)	OSADA SUCCESS 40 (OS-40)
Medical device approval number	23B2X00013
Classification	General medical device
Power input	Voltage: AC 230 V Frequency: 50/60 Hz Current: 0.1A (MAX: 1.25A)
Rotation speed	1,000 - 40,000 min <sup>-1</sup>
Body dimensions	W100×D185×H105mm
Body weight	Unit: 2.9 kg., Motor & Handpiece: 175 g

\* Color may vary from the image shown due to computer display.

\* The product specification is subject to change due to modification, without notice.

\*